
OSSEOCONDUCT βTCP



- 3rd Generation Beta Tricalcium Phosphate Granules For Bone Grafting
- βTCP Air Abrasion for Dental Appliances

STORAGE

If the gray temperature indicator located on the product packaging is **black** upon arrival, the efficacy of the graft material may have been compromised. Please contact SteinerBio immediately for instructions to replace the products: (866)317-1348



NOTIFY US IMMEDIATELY IF TEMPERATURE DOT IS BLACK

DIRECTIONS

OsseoConduct Granules

- Open the pouch and remove sterile packet containing the OsseoConduct Granules. Both the inside and outside of the packet is sterile and can be placed on the sterile operating surface.
- Open packet using sterile scissors and place the granules in an appropriate sterile container.
- To facilitate handing, the OsseoConduct Granules can be wetted with sterile water, autogenous blood, bone marrow, or a wetting agent of your choice.
- OsseoConduct Granules maintain graft site volume while osteoinduction occurs.

OsseoConduct Micron

- Open the pouch and remove sterile packet containing OsseoConduct Micron powder. Both the inside and outside of the packet is sterile and can be placed on the sterile operating surface.
- Open packet using sterile scissors and pour the powder into the air gun.
- Follow the air gun manufacturers directions for loading and operation.
- Surgically expose the implant surface and thoroughly surround the implant with a wet sterile cotton gauze.
- Mechanically remove any cement or calculus from the implant surface.
- Point the air gun nozzle directly toward the implant surface only. Never direct the air gun into soft connective tissues.
- Collision of the particles with the implant surface may cause minor sparks. Do not use this material on a patient breathing a high concentration of oxygen.
- When the implant is visibly clean, thoroughly rinse the entire surgical site with sterile water.

PRECAUTIONS

This product has not been studied in pediatric patients. Do not mix with any other graft materials other than SteinerBio bone grafts. Do not use for immediate support of removable dental appliances.

Federal law restricts this device to sale by or on the order of a dentist.

OSSEOCONDUCT βTCP

DESCRIPTION

OsseoConduct βTCP is a synthetic, porous, resorbable biocompatible beta tricalcium phosphate material.

OsseoConduct is available in two forms: a granular form for bone grafting and in an air abrasive form for cleaning implant surfaces.

This product contains no human or animal tissue.

INDICATIONS

OsseoConduct Granules are intended for use as a maxillofacial bone grafting material in bone voids and for appositional bone growth. OsseoConduct granules are designed to maintain shape and volume when used for ridge augmentation or socket grafting.

CONTRAINDICATIONS

OsseoConduct is not indicated where loading of the graft site is planned such as oral sites that are intended to be loaded immediately with dental prosthesis. This product is contraindicated in individuals who are unable to follow post-op instructions.

WARNINGS

OsseoConduct is single use only. Do not re-sterilize or reuse.
Do not use OsseoConduct if the package has been opened or damaged, or has exceeded the expiration date. Do not use if the temperature indicator on the packaging has turned black.

STEINERBIO
1051 Olsen St., Bldg 3611
Henderson, NV 89011
contactus@steinerbio.com
www.steinerbio.com
866.317.1348