RIDGE GRAFT KIT




REFRIGERATE PRODUCT UPON ARRIVAL

- Designed to be mixed chairside

STORAGE

Product will be shipped in an insulated shipping box and/or with cold packs, which may thaw during shipment. If the gray temperature indicator located on the product packaging is **black** upon arrival, the efficacy of the graft material may have been compromised. Please contact SteinerBio immediately for instructions to replace the products: (866)317-1348



NOTIFY US IMMEDIATELY IF TEMPERATURE DOT IS BLACK

DIRECTIONS

- It is very important not to compress the powder syringe before mixing.
- Hold both syringes upright and unscrew caps.
- Place the female barb connector to either syringe.
- Hold the powder syringe upright and screw both syringes together. Liquid will not leak.
- Slowly drip the liquid syringe into powder syringe and shake for a few seconds to mix.
- Using the syringe plungers, push them back and forth about ten (10) times until completely mixed.
- After mixing, push all of the mixed graft material into either syringe and unscrew both syringes.
- Dispense the Ridge Graft material into a sterile container, add the OsseoConduct βTCP granules, and mix together.
- Apply mixed graft material to site immediately not allowing graft material to dry.

PRECAUTIONS

Do not overfill defects. Do not compromise blood supply to the defect area. This product has not been studied in pediatric patients. Immediate Graft is not intended for immediate load bearing. Do not mix with any other graft materials other than SteinerBio bone grafts.

Federal law restricts this device to sale by or on the order of a dentist.

RIDGE GRAFT KIT

DESCRIPTION

Ridge Graft Kit is a combination of Ridge Graft and OsseoConduct βTCP Standard granules (500-1000µm). Ridge Graft is a dual phase calcium phosphate based synthetic bone graft material. Each syringe of Ridge Graft contains 1cc of graft material. Each packet of OsseoConduct βTCP contains 1cc of Standard granules. Ridge Graft is intended to be mixed with OsseoConduct βTCP Standard granules in a 1:1 ratio.

This product contains no human or animal tissue.

INDICATIONS

Ridge Graft Kit is a bone grafting material intended for use in ridge augmentation and sinus elevations.

CONTRAINDICATIONS

This product is contraindicated in individuals who are unable to follow post-op instructions.

WARNINGS

Ridge Graft Kit is single use only. Do not re-sterilize or reuse.
Do not use Ridge Graft Kit if the package has been opened or damaged, or has exceeded the expiration date. Do not use if the temperature indicator on the packaging has turned black.

STEINERBIO
1051 Olsen St., Bldg 3611
Henderson, NV 89011
contactus@steinerbio.com
www.steinerbio.com
866.317.1348