SINUS GRAFT




REFRIGERATE PRODUCT UPON ARRIVAL

- Designed to be mixed chairside
- Can be mixed 30 minutes before use

STORAGE

Product will be shipped in an insulated shipping box and/or with cold packs, which may thaw during shipment. If the gray temperature indicator located on the product packaging is **black** upon arrival, the efficacy of the graft material may have been compromised. Please contact SteinerBio immediately for instructions to replace the products: (866)317-1348



NOTIFY US IMMEDIATELY IF TEMPERATURE DOT IS BLACK

DIRECTIONS

- It is very important not to compress the powder syringe before mixing.
- Hold both syringes upright and unscrew caps.
- Place the female barb connector to either syringe.
- Hold the powder syringe upright and screw both syringes together. Liquid will not leak.
- Slowly drip the liquid syringe into powder syringe and shake for a few seconds to mix.
- Using the syringe plungers, push them back and forth about ten (10) times until completely mixed.
- After mixing, push all of the mixed graft material into yellow syringe and unscrew both syringes.
- Place the dispensing tip aside for 3 minutes for graft activation to occur.

PRECAUTIONS

Sinus Graft is specifically designed for hydraulically raising the maxillary sinus membrane for sinus augmentation. Do not compromise blood supply to the defect area. This product has not been studied in pediatric patients. Do not mix with any other graft materials other than SteinerBio bone grafts.

Federal law restricts this device to sale by or on the order of a dentist.

SINUS GRAFT

DESCRIPTION

Sinus Graft is a dual phase calcium phosphate based synthetic bone graft material. This product is designed to be quickly and completely resorbed, and to produce normal, healthy vital bone. This product contains no human or animal tissue.

INDICATIONS

Sinus Graft is a bone grafting material specifically designed for hydraulically raising the maxillary sinus membrane for sinus augmentation.

The primary indication is injection through a microsurgical osteotomy for hydraulically lifting the sinus membrane for either crestal or lateral lifts. Due to its flowable and low viscosity characteristics of use, it is not advised to be used with the classic lateral window sinus elevation.

For the large classic window elevation, we recommend using Ridge Graft Kit.

CONTRAINDICATIONS

This product is contraindicated in individuals who are unable to follow post-op instructions.

WARNINGS

Sinus Graft is single use only. Do not re-sterilize or reuse.
Do not use Sinus Graft if the package has been opened or damaged, or has exceeded the expiration date. Do not use if the temperature indicator on the packaging has turned black.

STEINERBIO
1051 Olsen St., Bldg 3611
Henderson, NV 89011
contactus@steinerbio.com
www.steinerbio.com
866.317.1348