SOCKET GRAFT INJECTABLE




REFRIGERATE PRODUCT UPON ARRIVAL

- Ready to use
- No mixing required

STORAGE

Product will be shipped in an insulated shipping box and/or with cold packs, which may thaw during shipment. If the gray temperature indicator located on the product packaging is **black** upon arrival, the efficacy of the graft material may have been compromised. Please contact SteinerBio immediately for instructions to replace the products: (866)317-1348



NOTIFY US IMMEDIATELY IF TEMPERATURE DOT IS BLACK

DIRECTIONS

- Keep the socket as dry as possible before injecting to avoid graft washout.
- Maintain the site as dry as possible to maintain control of the graft material.
- If significant bleeding occurs before grafting, dry the socket with a sterile 2x2 gauze wetted with 1:50 lidocaine. Wait for bleeding to slow, then quickly fill the socket with Socket Graft Injectable.
- Remove granulation tissue and any foreign material after extraction to avoid graft failure.
- Prepare the d-PTFE membrane before injecting by creating a full thickness flap (buccal and lingual) approximately 6mm and place membrane under lingual flap.
- Inject into the apex, fill toward the gingival margin, and place the membrane under the buccal flap.
- Suture the site or use the sutureless membrane technique by using Oral Bond.
- If you choose to temporize, use an essex retainer rather than a flipper.

SUTURELESS MEMBRANE TECHNIQUE

PRECAUTIONS

Do not overfill. Do not compromise blood supply to the defect area. This product has not been studied in pediatric patients. Socket Graft Injectable is not intended for immediate load bearing. Do not mix with any other graft materials other than SteinerBio bone grafts. Do not use animal or human membranes for sutures to cover Socket Graft Injectable or the graft may fail.

Federal law restricts this device to sale by or on the order of a dentist.

SOCKET GRAFT INJECTABLE

DESCRIPTION

Socket Graft Injectable is a dual phase calcium phosphate based synthetic bone graft material. It is designed to quickly and safely treat the extraction socket, to retain alveolar bone, stimulate bone formation, and reduce post-operative complications. This product is designed to be quickly and completely resorbed, and to produce normal, healthy vital bone. This product contains no human or animal tissue.

Ridge preservation insurance code: D7953

INDICATIONS

Socket Graft Injectable is a bone grafting material intended for use in dental extraction sockets that have all walls remaining. This product is cleared by the FDA for implant placement at 8 weeks.

CONTRAINDICATIONS

This product is contraindicated in individuals who are unable to follow post-op instructions.

WARNINGS

Socket Graft Injectable is single use only. Do not re-sterilize or reuse.
Do not use Socket Graft Injectable if the package has been opened or damaged, or has exceeded the expiration date. Do not use if the temperature indicator on the packaging has turned black.

STEINERBIO
1051 Olsen St., Bldg 3611
Henderson, NV 89011
contactus@steinerbio.com
www.steinerbio.com
866.317.1348